

FIG. 2 (PRIOR ART)






FIG. 3






## DIFFERENT METHODS OF CALIBRATION FILE=030AMCG.ALL; 169 SAMPLES

**CALIBRATION** 

**CROSS VALIDATION** 




"A"=1ST ORDER (14 OD); "B"=1ST+2ND ORDER
"C"=1+2+3 ORDER; "D"="A"+PULSE+Td;
"E"=1+2+P+Td; "F"="B"+P+Td J121B

FIG. 4




## SEC VS. SECV USING THE 10 BEST CALIBRATIONS USING 1ST, 2ND, AND 3RD ORDER TERMS



J120 USING CAL FILE 039amcg.all. USED MONTE CARLO WITH 1,000,000 RANDOM WL REGRESSIONS. (ANALYSIS DONE 9/99)

FIG.5





USING SIX DBAs CHOSEN AT RANDOM

COEFFICIENT OF DETERMINATION (R × R)

0.9

0.8

0.7

0.6

0.5

0.4

0.3

0.2

DIFFERENT CALIBRATION APPROACHES

THE PROTOCOL THAT PASSED THE 4 UNITS CRITERIA = 0.642 × -AVERAGE GLUCOWATCH (FDA SUBMITTAL

FIG. 6 FOR UNITS 1002, 1012, 1020, 1021, 3013
AND 3015. 20 VARIABLES WERE USED FOR ALL
ANALYSES EXCEPT FOR THE 14WL. FILE: J400

4 OK DBAs

3 INDEX

Lo->Hi-> Mi-> Bsc

 $2 \times INDEX$ 

Physical

2&3 ORDERS

14 WL

0.1

0

**AVERAGE** 

1 SIGMA LIMITS